

EC - DECLARATION OF CONFORMITY



CE – DECLARATION DE CONFORMITE
EG – ÜBEREINSTIMMUNGSERKLÄRUNG
CE – DICHIARAZIONE DI CONFORMITÀ
CE – DECLARACION DE CONFORMIDAD

MODEL:

Modèle, Modell, Modello, Modelo

SL3-1, SL3-SDI-1, SL3-XMTR-1, WSDI-M/S, X-Link-1
CDMA Link-1, GPRS Link-1, HSPA Link-1, IRIDIUM Link-1

Designation:

Beschreibung, Designación, Désignation, Denominazione

Satlink3 & XLink Loggers/Transmitters & Wireless SDI-12

THIS PRODUCT COMPLIES WITH THE FOLLOWING EUROPEAN COMMUNITY DIRECTIVES

Ce produit est conforme aux directives de la Communauté Européenne suivantes
Dieses Produkt entspricht den nachstehend aufgeführten Richtlinien der Europäischen Union
Questo prodotto è conforme alle seguenti direttive della Comunità Europea
Este producto cumple con las directivas siguientes de la Comunidad Europea

2004/108/EC (until April 19 th , 2016) 2014/30/EU (from April 20 th , 2016)	EMC Directive
1999/5/EC (until June 12 th , 2016) 2014/53/EU (from June 13 th , 2016)	Radio Equipment & Telecommunications Terminal Equipment Radio Equipment Directive
2006/95/EC (until April 19 th , 2016) 2014/35/EU (from April 20 th , 2016)	Low Voltage Directive
2011/65/EU (from July 21 st , 2017)	Restriction of Hazardous Substances (RoHS 2)

THE FOLLOWING STANDARDS WERE USED TO VERIFY COMPLIANCE WITH THE DIRECTIVES

Les normes suivantes ont été appliquées pour vérifier que ce produit soit conforme aux directives
Die folgenden Normen wurden zur Prüfung der Übereinstimmung mit diesen Richtlinien angewendet
Sona state usate lo seguenti norme per verificare la conformità alle direttive
Las normas siguientes han sido utilizadas para verificar que el producto cumpla con las directivas

EN 61326-1: 2013	Electrical Equipment for Measurement, Control & Lab Use
EN 301 489-1 V1.9.2	Radio Spectrum Matters: Part 1
EN 301 489-3 V1.6.1	Radio Spectrum Matters: Part 3
EN 61010-1: 2010	Safety Requirements for EE for Measurement, Control & Lab Use
EN 50581:2012	Guiding Standard for Compliance with RoHS2 Requirements

**THIS PRODUCT WAS TESTED BY AN ISO 17025 ACCREDITED LABORATORY AND WAS
MANUFACTURED BY AN ISO 9001 ACCREDITED FACILITY.**

Year CE Mark Originally Affixed to Product: 2016

Année de l'application originale du marquage CE au produit
Jahr der ersten CE Markierung des Produktes
Anno del marchio CE iniziale del prodotto
Año de la primera aplicación de la marca CE al producto



Dan Farrell
Chief Technology Officer (CTO)

19 Feb., 2016

Date
Date, Datum, Data, Fecha