

Highly Accurate, Solid State Pressure Transducer

# Accubar<sup>®</sup> Barometric Pressure Sensor 5600-0120



## Specifications

Specifications subject to change without notice

RANGE	500 to 1100mB (hPa)	
	14.8 to 32.5 in. Hg	
ELEVATION	-2,300 to 18,300 feet (-700 to 5,600 meters)	
ACCURACY	0.5mB (hPa) over temperature range Consult factory for greater accuracy.	
RESOLUTION	0.01mB	
LONG TERM STABILITY	±0.1 mB (hPa)/year	
AVERAGE TIME	0 to 240 seconds (user selectable)	
TEMP RANGE	-40°C to +60°C	
SUPPLY POWER	+8 to +30 VDC (reverse polarity protected)	
OUTPUT	Standard (-1)	RS-232 & SDI-12
	Analog (-2)	0-5VDC (programmable range)
RS-232 DATA	Continuous or on-demand	
COMMUNICATION	1200 baud, even parity, 7 data bits, one stop bit	
PRESSURE UNITS	mB, hPa, kPa, in. Hg, mm Hg, ATM, psi, user specified units	
DIMENSIONS	3.14 in. x 2.95 in. x 2.24 in. excluding connectors and cable	
WEIGHT	<1 lb	
VIBRATION	0.5g @ 10-50Hz	
PRESSURE FITTING	Hose barb for 3/16 in. ID hose	
ELECTRICAL	Cable gland for 1/4 in. cable to	
CONNECTION	Terminal block for 16 to 28 AWG wire	
POWER CONSUMPTION 5600-0120-1	RS-232 active	17 mA typ
	SDI-12 active	6 mA typ
	standby	0.2 mA typ
POWER CONSUMPTION 5600-0120-2	RS-232 active	18 mA typ
	SDI-12 active	7 mA typ
	standby	0.8 mA typ

## Features

- ▶ Low power design makes it ideally suited for remote monitoring applications.
- ▶ Capable of operating from elevations of 2300 feet below sea level to 18,300 feet above sea level
- ▶ Wide temp range -40°C to +60°C
- ▶ Low power consumption
- ▶ Solid state construction
- ▶ Low cost
- ▶ Light weight aluminum construction
- ▶ RS-232, SDI-12, and Analog 0 - 5 VDC outputs

## Ordering

5600-0120-1

Pressure Sensor, Barometric, RS-232 and SDI-12 outputs

5600-0120-2

Pressure Sensor, Barometric, RS-232, SDI-12 and Analog 0-5VDC outputs



**SUTRON CORPORATION**

22400 Davis Drive  
GS-25F604D SBSA

Sterling, VA 20164  
NESDIS Certified

(703) 406-2800  
ISO Certified

(703) 406-2801 Fax  
EUMETSAT Certified

www.sutron.com  
CE Certified

sales@sutron.com  
INSAT Certified

page 1  
9/8/2011