

Highly Accurate, Solid State Pressure Transducer

# Accubar<sup>®</sup> Barometric Pressure Sensor 5600-0120



Specifications	
Specifications subject to change without notice	
RANGE	500 to 1100mB (hPa)
	14.8 to 32.5 in. Hg
ELEVATION	-2,300 to 18,300 feet (-700 to 5,600 meters)
ACCURACY	0.5mB (hPa) over temperature range Consult factory for greater accuracy.
RESOLUTION	0.01mB
LONG TERM STABILITY	±0.1 mB (hPa)/year
AVERAGE TIME	0 to 240 seconds (user selectable)
TEMP RANGE	-40°C to +60°C
SUPPLY POWER	+8 to +30 VDC (reverse polarity protected)
OUTPUT	Standard (-1)      RS-232 & SDI-12
	Analog (-2)      0-5VDC (programmable range)
RS-232 DATA	Continuous or on-demand
COMMUNICATION	1200 baud, even parity, 7 data bits, one stop bit
PRESSURE UNITS	mB, hPa, kPa, in. Hg, mm Hg, ATM, psi, user specified units
DIMENSIONS	3.14 in. x 2.95 in. x 2.24 in. excluding connectors and cable
WEIGHT	<1 lb
VIBRATION	0.5g @ 10-50Hz
PRESSURE FITTING	Hose barb for 3/16 in. ID hose
ELECTRICAL	Cable gland for 1/4 in. cable to
CONNECTION	Terminal block for 16 to 28 AWG wire
POWER CONSUMPTION 5600-0120-1	RS-232 active 17 mA typ SDI-12 active 6 mA typ standby 0.2 mA typ
POWER CONSUMPTION 5600-0120-2	RS-232 active 18 mA typ SDI-12 active 7 mA typ standby 0.8 mA typ

## Features

- ▶ Low power design makes it ideally suited for remote monitoring applications.
- ▶ Capable of operating from elevations of 2300 feet below sea level to 18,300 feet above sea level
- ▶ Wide temp range -40°C to +60°C
- ▶ Low power consumption
- ▶ Solid state construction
- ▶ Low cost
- ▶ Light weight aluminum construction
- ▶ RS-232, SDI-12, and Analog 0 - 5 VDC outputs

## Ordering

5600-0120-1

Pressure Sensor, Barometric, RS-232 and SDI-12 outputs

5600-0120-2

Pressure Sensor, Barometric, RS-232, SDI-12 and Analog 0-5VDC outputs



**SUTRON CORPORATION**

22400 Davis Drive  
GS-25F604D SBSA

Sterling, VA 20164  
NESDIS Certified

(703) 406-2800  
ISO Certified

(703) 406-2801 Fax  
EUMETSAT Certified

www.sutron.com  
CE Certified

sales@sutron.com  
INSAT Certified

page 1  
9/8/2011